

SKLÁRNÝ MORAVIA, akciová společnost
Úsobrno č.p. 79
679 39 Úsobrno
CZECH REPUBLIC

Tel.: + 420 516 427 711
Fax: + 420 516 427 700
E-mail: info@sklomoravia.cz

Declaration of Conformity

Product: Laboratory glass containers
Capacity: 50 – 2000 ml
Color: Flint, Amber

All glass containers manufactured by Sklarny Moravia have been manufactured in conformity with drawing, product specification and our standard procedures.

All laboratory glass containers are manufactured in conformity with:

- ISO 719:1985
- 94/62/EC
- EC 1935/2004 *
- EC 2023/2006 *

They are also manufactured in conformity with standards referred to certificate No. CQ-32/2019, issued by IKATES s.r.o. Teplice, certificate authority No. 3065.

Company quality management system is certified according to ISO 9001:2015 by Intertek.

* Laboratory glass containers are intended for storage of chemical substances and mixtures. We recommend rinsing the containers with distilled water before use. In the case of grinded containers, there may occur glass dust on the sides and bottom of the bottle from grinding process. Such containers must be rinsed before use. We recommend rinsing with a 2% acetic acid solution followed by distilled water. If containers are intended for food and beverage storage, the beverage / food manufacturer must rinse the containers properly with drinking water or water with low % alcohol before filling! Grinded containers and stoppers are recommended to be rinsed first with 2% acetic acid solution followed by drinking water or water with low % alcohol!

Absolute tightness cannot be guaranteed for grinded containers with grinded stoppers. Therefore we do not recommend to transport liquids. Any use of waxes and sealing material to ensure better tightness is solely the responsibility of the user of the sealing material.

Úsobrno, January 2021

Issued by: Martin Dudek